

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.60 μm	50.40 μm	-0.20 μm	± 2 μm	±	5.8E-01 μm
120.30 μm	119.6 μm	-0.70 μm	± 2 μm	±	5.8E-01 μm
246.20 μm	246.0 μm	-0.20 μm	± 2 μm	±	5.8E-01 μm
479.00 μm	478.0 μm	-1.00 μm	± 2 μm	±	5.8E-01 μm