

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
52.65 μm	52.70 μm	0.05 μm	$\pm 0.1 \mu\text{m}$	$\pm 5.8\text{E-}02 \mu\text{m}$
155.23 μm	155.5 μm	0.27 μm	$\pm 1 \mu\text{m}$	$\pm 6.2\text{E-}02 \mu\text{m}$
257.12 μm	257.3 μm	0.18 μm	$\pm 1 \mu\text{m}$	$\pm 7.0\text{E-}02 \mu\text{m}$
511.71 μm	512.0 μm	0.29 μm	$\pm 1 \mu\text{m}$	$\pm 9.7\text{E-}02 \mu\text{m}$