

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49922	0.780	± 0.8	±	4.3E-01 mm						
0 a 1	0.99924	0.760	± 1	±	4.3E-01 mm						
0 a 1.5	1.49917	0.830	± 1.2	±	4.8E-01 mm						
0 a 2	1.99925	0.750	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49923	0.770	± 1.6	±	4.8E-01 mm						
0 a 3	2.99919	0.810	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49924	0.760	± 2	±	4.8E-01 mm						
0 a 4	3.99917	0.830	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49851	1.490	± 2.4	±	4.8E-01 mm						
0 a 5	4.99930	0.700	± 2.6	±	4.8E-01 mm						