

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	28.4 °C	-1.662 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	0.0 %HR	-30.161 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.376 %HR	20.0 %HR	-30.376 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	35.0 %HR	-35.652 %HR	± 5.0 %HR	± 1.9E+00 %HR