Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
2.00 in	2.0000 in	0 in	$\pm 0.003$ in	± 7.0E-02 in	
2.05 in	2.0500 in	0 in	$\pm 0.003$ in	± 7.6E-02 in	
2.10 in	2.1000 in	0 in	$\pm 0.003$ in	± 9.4E-02 in	
2.15 in	2.1500 in	0 in	$\pm 0.003$ in	± 7.6E-02 in	
2.20 in	2.2000 in	0 in	$\pm 0.003$ in	± 9.4E-02 in	
2.25 in	2.2500 in	0 in	$\pm 0.003$ in	± 6.9E-02 in	
2.30 in	2.3000 in	0 in	$\pm 0.003$ in	± 1.1E-01 in	
2.35 in	2.3500 in	0 in	$\pm 0.003$ in	± 1.3E-01 in	
2.40 in	2.4000 in	0 in	$\pm 0.003$ in	± 1.4E-01 in	