

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
5.0135 ml	5 ml	-0.014 ml	± 0.1 ml	±	1.8E-01 µl
50.1139 ml	50 ml	-0.114 ml	± 0.1 ml	±	1.6E+00 µl
100.1731 ml	100 ml	-0.173 ml	± 0.1 ml	±	3.2E+00 µl