

Certificado No. (Certificate #): 59052		Equipo (Instrument): Trazador de Alturas			
Fecha de Calibración (Calibration Date): 2025-07-16		Marca (Brand): No Indcada			
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Altura (Height)	1.000 in	1.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	2.000 in	2.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	4.000 in	4.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	5.000 in	5.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	6.000 in	6.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	7.000 in	7.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	9.000 in	9.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	10.000 in	10.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	13.000 in	13.000 in	0.0000 in		± 7.2E-02 in
Altura (Height)	15.000 in	15.000 in	0.0000 in		± 7.2E-02 in