

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.306 °C	2.5 °C	0.194 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.114 °C	5.3 °C	0.186 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.686 °C	9.8 °C	0.114 °C	± 0.5 °C	± 4.0E-01 °C