

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.306 °C	2.6 °C	0.294 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.114 °C	5.5 °C	0.386 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.686 °C	10.1 °C	0.414 °C	± 0.5 °C	± 4.0E-01 °C