

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 2 °C              | 2.306 °C          | 2.5 °C    | 0.194 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.114 °C          | 5.4 °C    | 0.286 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 9.686 °C          | 9.8 °C    | 0.114 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |