

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0010 in	0.001 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0020 in	0.002 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0030 in	0.003 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0040 in	0.004 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0050 in	0.005 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0060 in	0.006 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0080 in	0.008 in	0.0000 in	± 0.002 in	±	2.9E-04 in
0.0100 in	0.010 in	0.0000 in	± 0.002 in	±	2.9E-04 in