

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
2 mm	1.94 mm	-60 µm	± 30 µm	±	5.8E+00 µm
5 mm	4.93 mm	-70 µm	± 30 µm	±	5.8E+00 µm
15 mm	15.03 mm	30 µm	± 30 µm	±	5.8E+00 µm
20 mm	20.13 mm	130 µm	± 30 µm	±	5.8E+00 µm
25 mm	25.26 mm	260 µm	± 30 µm	±	5.8E+00 µm