

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 5	4.99909	0.011	± 0.26	± 1.4E-03 in					
0 a 10	9.99946	0.007	± 0.26	± 1.4E-03 in					
0 a 15	14.99955	0.005	± 0.26	± 1.4E-03 in					
0 a 20	19.99965	0.004	± 0.26	± 1.4E-03 in					
0 a 25	24.99946	0.007	± 0.26	± 1.4E-03 in					
0 a 30	29.99913	0.010	± 0.26	± 1.4E-03 in					
0 a 35	34.99924	0.009	± 0.26	± 1.4E-03 in					
0 a 40	39.99941	0.007	± 0.26	± 1.4E-03 in					
0 a 45	44.99945	0.007	± 0.26	± 1.4E-03 in					
0 a 50	49.99925	0.009	± 0.26	± 1.4E-03 in					