

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
23.3 μm	24.5 μm	1.2 μm	$\pm 2.5 \mu\text{m}$	$\pm 5.8\text{E-}02 \mu\text{m}$
51.6 μm	53.8 μm	2.2 μm	$\pm 2.5 \mu\text{m}$	$\pm 5.8\text{E-}02 \mu\text{m}$
124.6 μm	128 μm	3.4 μm	$\pm 3.7 \mu\text{m}$	$\pm 6.1\text{E-}02 \mu\text{m}$
250.3 μm	257 μm	6.7 μm	$\pm 7.5 \mu\text{m}$	$\pm 6.9\text{E-}02 \mu\text{m}$
486 μm	498 μm	12.0 μm	$\pm 14.6 \mu\text{m}$	$\pm 9.3\text{E-}02 \mu\text{m}$
963 μm	985 μm	22.0 μm	$\pm 28.9 \mu\text{m}$	$\pm 1.6\text{E-}01 \mu\text{m}$