

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.012 °C	29.8 °C	-0.212 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.9 %HR	1.330 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.631 %HR	51.4 %HR	0.769 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	70.7 %HR	0.355 %HR	± 3.0 %HR	± 1.9E+00 %HR