

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.012 °C	30.0 °C	-0.012 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.0 %HR	1.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.631 %HR	51.7 %HR	1.069 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	69.9 %HR	-0.445 %HR	± 3.0 %HR	± 1.9E+00 %HR