

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
17 lx	17.51 lx	0.51 lx	± 0.85 lx	±	9.9E-02 lx
50 lx	50.78 lx	0.78 lx	± 2.5 lx	±	9.9E-02 lx
100 lx	101.21 lx	1.21 lx	± 5.0 lx	±	9.9E-02 lx
198 lx	199.91 lx	1.91 lx	± 9.9 lx	±	9.9E-02 lx
303 lx	299.38 lx	-3.62 lx	± 15 lx	±	6.4E-01 lx
503 lx	501.33 lx	-1.67 lx	± 25 lx	±	6.4E-01 lx
714 lx	709.83 lx	-4.17 lx	± 36 lx	±	8.9E-01 lx
1006 lx	998.81 lx	-7.19 lx	± 50 lx	±	2.9E+00 lx
2000 lx	1986.61 lx	-13.39 lx	± 100 lx	±	2.9E+00 lx