

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
193.2 lx	195.2009 lx	-1.9808 lx	± 20 lx	±	4.5E+00 lx
340.7 lx	337.1100 lx	3.6378 lx	± 30 lx	±	3.0E+00 lx
754.5 lx	749.6364 lx	4.8719 lx	± 50 lx	±	1.0E+00 lx
926.6 lx	919.9091 lx	6.6521 lx	± 60 lx	±	1.6E+01 lx
2,878.7 lx	2,859.8182 lx	18.8855 lx	± 160 lx	±	7.8E+01 lx
51,091.6 lx	51,001.3636 lx	90.2058 lx	± 2600 lx	±	9.1E+02 lx