

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
193.2 lx	195.20 lx	-1.98 lx	± 10 lx	±	9.9E-02 lx
340.7 lx	337.11 lx	3.64 lx	± 17 lx	±	6.4E-01 lx
754.5 lx	749.64 lx	4.87 lx	± 38 lx	±	1.2E+00 lx
926.6 lx	919.91 lx	6.65 lx	± 46 lx	±	1.2E+00 lx
2,878.7 lx	2,859.82 lx	18.89 lx	± 144 lx	±	7.9E+00 lx
51,091.6 lx	51,001.36 lx	90.21 lx	± 2555 lx	±	4.3E+02 lx