

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0116 ml	10 ml	-0.012 ml	± 0.02 ml	±	3.2E-01 µl
24.9639 ml	25 ml	0.036 ml	± 0.05 ml	±	8.0E-01 µl
49.9829 ml	50 ml	0.017 ml	± 0.05 ml	±	1.6E+00 µl