

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1,000.3019 ml	1000 ml	-0.302 ml	± 0.3 ml	±	3.5E-01 µl