

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0045 ml	1 ml	-0.005 ml	± 0.5 ml	±	9.4E-02 µl
4.9966 ml	5 ml	0.003 ml	± 0.5 ml	±	2.9E-01 µl
9.9973 ml	10 ml	0.003 ml	± 0.5 ml	±	5.7E-01 µl