

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.514 °C	29.9 °C	-0.614 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.018 %HR	33 %HR	2.982 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.599 %HR	48 %HR	-1.599 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	73.832 %HR	73 %HR	-0.832 %HR	± 5.0 %HR	± 1.9E+00 %HR