

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.174 s	3601.00 s	0.826 s	0.023%	± 0.005 %	± 6.3E-02 s
3600.547 s	3601.00 s	0.453 s	0.013%	± 0.005 %	± 6.3E-02 s
3600.395 s	3601.00 s	0.605 s	0.017%	± 0.005 %	± 6.3E-02 s