

Certificado No. (Certificate #): 59145

Equipo (Instrument): Medidor de Espesores

Fecha de Calibración (Calibration Date): 2025-07-23

Marca (Brand): -

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.0000 mm	1.003 mm	3.0000 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
3.0000 mm	3.002 mm	2.0000 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
6.0000 mm	6.001 mm	0.5000 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
9.0000 mm	8.997 mm	-2.7500 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
12.0000 mm	11.995 mm	-4.7500 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
15.0000 mm	14.992 mm	-7.7500 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
18.0000 mm	17.994 mm	-5.7500 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
20.0000 mm	19.992 mm	-8.5000 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
25.0000 mm	24.996 mm	-4.0000 μm	$\pm \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$