

Certificado No. (Certificate #):		59145		Equipo (Instrument):		Medidor de Espesores			
Fecha de Calibración (Calibration Date):		2025-07-23		Marca (Brand):		-			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
1.000 mm		1.003 mm		3 µm		± µm		±	5.8E-01 µm
3.000 mm		3.002 mm		2 µm		± µm		±	5.8E-01 µm
6.000 mm		6.001 mm		1 µm		± µm		±	5.8E-01 µm
9.000 mm		8.997 mm		-3 µm		± µm		±	5.8E-01 µm
12.000 mm		11.995 mm		-5 µm		± µm		±	5.8E-01 µm
15.000 mm		14.992 mm		-8 µm		± µm		±	5.8E-01 µm
18.000 mm		17.994 mm		-6 µm		± µm		±	5.8E-01 µm
20.000 mm		19.992 mm		-8 µm		± µm		±	5.8E-01 µm
25.000 mm		24.996 mm		-4 µm		± µm		±	5.8E-01 µm