

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.180 °C	0.1 °C	-0.080 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.0 °C	0.007 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.238 °C	10.0 °C	-0.238 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	32.6 %HR	2.337 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	52.6 %HR	1.917 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	71.2 %HR	0.957 %HR	± 3.0 %HR	± 1.9E+00 %HR