

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.223 °C	12.5 °C	0.277 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.084 °C	27.2 °C	0.116 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.657 °C	42.6 °C	-0.057 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.656 %HR	53.8 %HR	1.144 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.032 %HR	72.0 %HR	-0.032 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.329 %HR	91.9 %HR	-0.429 %HR	± 3.0 %HR	± 1.9E+00 %HR