

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.180 °C	0.5 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.4 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.238 °C	10.5 °C	0.262 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	30.5 %HR	0.237 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	51.4 %HR	0.717 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.7 %HR	-0.543 %HR	± 3.0 %HR	± 1.9E+00 %HR