

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.180 °C	0.5 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.4 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.238 °C	10.4 °C	0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	32.1 %HR	1.837 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	52.8 %HR	2.117 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.1 %HR	1.857 %HR	± 3.0 %HR	± 1.9E+00 %HR