

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | 0.180 °C                   | 0.4 °C             | 0.220 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 5 °C                       | 4.993 °C                   | 5.2 °C             | 0.207 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 10 °C                      | 10.238 °C                  | 10.2 °C            | -0.038 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.263 %HR                 | 29.6 %HR           | -0.663 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.683 %HR                 | 52.3 %HR           | 1.617 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.243 %HR                 | 69.9 %HR           | -0.343 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |