

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.223 °C	12.5 °C	0.277 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.084 °C	27.0 °C	-0.084 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.657 °C	42.5 °C	-0.157 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.656 %HR	53.5 %HR	0.844 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.032 %HR	74.5 %HR	2.468 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.329 %HR	91.2 %HR	-1.129 %HR	± 3.0 %HR	± 1.9E+00 %HR