

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.223 °C	12.3 °C	0.077 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.084 °C	27.5 °C	0.416 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.657 °C	42.8 °C	0.143 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.656 %HR	55.0 %HR	2.344 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.032 %HR	73.9 %HR	1.868 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.329 %HR	92.3 %HR	-0.029 %HR	± 3.0 %HR	± 1.9E+00 %HR