

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.514 °C	29.8 °C	-0.714 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.018 %HR	32 %HR	1.982 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.599 %HR	47 %HR	-2.599 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	73.832 %HR	73 %HR	-0.832 %HR	± 5.0 %HR	± 1.9E+00 %HR