

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.351 °C	-0.3 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.536 °C	20.4 °C	-0.136 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.176 °C	34.8 °C	-0.376 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	33.8 %HR	3.537 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	50.038 %HR	52.5 %HR	2.462 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	79.5 %HR	-0.661 %HR	± 3.0 %HR	± 1.9E+00 %HR