

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.351 °C	-0.6 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.4 °C	0.365 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.176 °C	34.9 °C	-0.276 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	33.2 %HR	2.937 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	50.038 %HR	51.5 %HR	1.462 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	78.5 %HR	-1.661 %HR	± 3.0 %HR	± 1.9E+00 %HR