

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 10.308 °C         | 10.5 °C   | 0.192 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 15 °C             | 15.021 °C         | 14.7 °C   | -0.321 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.610 °C         | 25.8 °C   | 0.190 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.263 %HR        | 32.1 %HR  | 1.837 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.243 %HR        | 70.1 %HR  | -0.143 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |