

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.308 °C	10.6 °C	0.292 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.4 °C	0.379 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.610 °C	25.8 °C	0.190 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	33.1 %HR	2.837 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	73.1 %HR	2.857 %HR	± 3.0 %HR	± 1.9E+00 %HR