

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60.071 s	60 s	-0.011 s	-0.0183%	± 0.0556%	±	2.0E-01 s
180.317 s	180 s	-0.037 s	-0.0205%	± 0.0556%	±	2.0E-01 s
300.253 s	300 s	-0.063 s	-0.0210%	± 0.0556%	±	2.0E-01 s
3600.146 s	3600 s	-0.146 s	-0.0041%	± 0.0556%	±	5.6E-02 s