

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1006 RPM	1000 RPM	-6 RPM	± 10 RPM	±	6.5E+01 RPM
3021 RPM	3000 RPM	-21 RPM	± 30 RPM	±	6.5E+01 RPM
3533 RPM	3500 RPM	-33 RPM	± 35 RPM	±	6.5E+01 RPM
4036 RPM	4000 RPM	-36 RPM	± 40 RPM	±	5.8E+01 RPM
4542 RPM	4500 RPM	-42 RPM	± 45 RPM	±	5.8E+01 RPM