

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
# 0 mm	0.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
1.0 mm	1.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.50 mm	1 µm	± 30 µm	±	5.8E+00 µm
2.0 mm	2.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
3 mm	3.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
5 mm	5.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
10 mm	10.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
15 mm	15.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
20 mm	20.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
25 mm	25.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
30 mm	30.01 mm	10 µm	± 30 µm	±	5.8E+00 µm