

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
2022 RPM	2000 RPM	-22 RPM	± 20 RPM	±	5.8E+02 RPM
5042 RPM	5000 RPM	-42 RPM	± 20 RPM	±	5.8E+02 RPM
6059 RPM	6000 RPM	-59 RPM	± 20 RPM	±	5.8E+02 RPM
8068 RPM	8000 RPM	-68 RPM	± 20 RPM	±	5.8E+01 RPM
10108 RPM	10000 RPM	-108 RPM	± 20 RPM	±	5.8E+01 RPM