

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.241 °C	-0.2 °C	0.041 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	14.9 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.0 °C	-0.363 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.376 %HR	9.3 %HR	-26.076 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.038 %HR	20.8 %HR	-29.238 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	39.5 %HR	-30.845 %HR	± 3.0 %HR	± 1.9E+00 %HR