

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.241 °C	0.0 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.0 °C	-0.363 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.376 %HR	33.3 %HR	-2.076 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.038 %HR	47.8 %HR	-2.238 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	67.4 %HR	-2.945 %HR	± 3.0 %HR	± 1.9E+00 %HR