

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.2 μm	24.1 μm	0.9 μm	± 2.5 μm	±	5.8E-02 μm
51.1 μm	51.8 μm	0.7 μm	± 2.5 μm	±	5.8E-02 μm
125.1 μm	126 μm	0.9 μm	± 3.8 μm	±	6.1E-02 μm
478.0 μm	478 μm	0.0 μm	± 14.3 μm	±	9.3E-02 μm
952 μm	961 μm	9.0 μm	± 28.6 μm	±	1.6E-01 μm