

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	23.4 μm	0.10 μm	± 3 μm	±	5.8E-01 μm
51.6 μm	51.5 μm	-0.10 μm	± 3 μm	±	5.8E-01 μm
124.6 μm	124.4 μm	-0.20 μm	± 3 μm	±	5.9E-01 μm
250.3 μm	250.0 μm	-0.30 μm	± 3 μm	±	6.3E-01 μm
486.00 μm	485.7 μm	-0.30 μm	± 3 μm	±	7.5E-01 μm
963.00 μm	963.5 μm	0.50 μm	± 3 μm	±	1.1E+00 μm