

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	23.0 μm	-0.30 μm	± 3 μm	±	5.8E-01 μm
30.00 μm	30.0 μm	0.00 μm	± 3 μm	±	5.8E-01 μm
46.50 μm	46.5 μm	0.00 μm	± 3 μm	±	5.8E-01 μm
50.00 μm	50.0 μm	0.00 μm	± 3 μm	±	5.8E-01 μm