

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.020 °C	0.2 °C	0.220 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.2 °C	-0.236 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.277 °C	35.6 °C	0.323 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	31.2 %HR	0.834 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.345 %HR	56.2 %HR	0.855 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.570 %HR	81.3 %HR	0.730 %HR	± 3.0 %HR	± 1.9E+00 %HR