

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.020 °C	0.0 °C	0.020 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.6 °C	0.164 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.277 °C	35.2 °C	-0.077 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	30.6 %HR	0.234 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.345 %HR	54.6 %HR	-0.745 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.570 %HR	80.2 %HR	-0.370 %HR	± 3.0 %HR	± 1.9E+00 %HR