

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 0 °C              | -0.020 °C         | 0.2 °C    | 0.220 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 20 °C             | 20.436 °C         | 20.4 °C   | -0.036 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 35 °C             | 35.277 °C         | 35.2 °C   | -0.077 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.366 %HR        | 31.7 %HR  | 1.334 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 55 %HR            | 55.345 %HR        | 56.5 %HR  | 1.155 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 80 %HR            | 80.570 %HR        | 78.1 %HR  | -2.470 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |